

# GC Application

ID No.: 10724

## FFAP Test Mix

**Column:** Zebtron<sup>™</sup> ZB-FFAP, GC Cap. Column 30 m x 0.32 mm x 0.25 μm, Ea

**Phase:** Nitroterephthalic Acid Modified Polyethylene Glycol

**Dimensions:** 30 meters x 0.32 mm x 0.25 μm

**Order No:** 7HM-G009-11

**Oven Profile:** 140°C Isothermal


**Carrier Gas:** Hydrogen, 1.93 mL/min

**Injection:** Split 50:1 1 μL @ 250°C

**Detection:** Flame Ionization (FID) (255°C)



Products used in this application:



## FFAP Test Mix

### ANALYTES:

- 1 Methane
- 2 Tridecane
- 3 Tetradecane
- 4 1-Octanol
- 5 Methyl decanoate
- 6 Isovaleric acid
- 7 Methyl undecanoate
- 8 Naphthalene
- 9 Methyl dodecanoate
- 10 1-Hexanoic acid
- 11 2,6-Dimethylphenol

