

GC Application

ID No.: 10724

FFAP Test Mix

Column: Zebtron™ ZB-FFAP, GC Cap. Column 30 m x 0.32 mm x 0.25 μm, Ea

Phase: Nitroterephthalic Acid Modified Polyethylene Glycol

Dimensions: 30 meters x 0.32 mm x 0.25 μm

Order No: 7HM-G009-11

Oven Profile: 140°C Isothermal


Carrier Gas: Hydrogen, 1.93 mL/min

Injection: Split 50:1 1 μL @ 250°C

Detection: Flame Ionization (FID) (255°C)



Products used in this application:



FFAP Test Mix

ANALYTES:

- 1 Methane
- 2 Tridecane
- 3 Tetradecane
- 4 1-Octanol
- 5 Methyl decanoate
- 6 Isovaleric acid
- 7 Methyl undecanoate
- 8 Naphthalene
- 9 Methyl dodecanoate
- 10 1-Hexanoic acid
- 11 2,6-Dimethylphenol

