

HPLC Application

ID No.: 11072

Increase Sample Through-put - Figure 2b

Column: Luna® 3 µm C18(2) 100 Å, LC Column 20 x 2 mm, Ea

Dimensions: 20 x 2 mm ID

Order No: 00M-4251-B0

Elution Type: Isocratic

Eluent A: Water w/0.1% ammonium hydroxide

Eluent B: Acetonitrile w/0.1% ammonium hydroxide

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	80	20

Flow Rate: 0.5 mL/min

Col. Temp.: 25 °C

Detection: UV-Vis Abs.-Variable Wave.(UV) @ 230 nm (ambient)


Analyst Note: SecurityGuard™ Guard Cartridge System extends column lifetime.

- SecurityGuard Cartridges, C18 4 x 2.0mm, 10/Pk Part No.: AJ0-4286

- Holder Part No.: KJ0-4282



Products used in this application:



ANALYTES:

1 Pindolol

