

HPLC Application

ID No.: 1303

Amino Acids

Column: Luna® 3 µm C18(2) 100 Å, LC Column 150 x 2 mm, Ea

Dimensions: 150 x 2 mm ID

Order No: 00F-4251-B0

Elution Type: Isocratic

Eluent A: 0.1 M Sodium Acetate, pH 3.8 w/Acetic Acid

Gradient Profile:	Step No.	Time (min)	Pct A
	1	0	100
	1	0	100
	2	14	100
	3	30	100

Flow Rate: 0.12 mL/min

Col. Temp.: 40 °C

Detection: Fluorescence (FLUOR) @ 385 nm (ambient)


Analyst Note: SecurityGuard™ Guard Cartridge System extends column lifetime.

- SecurityGuard Cartridges, C18 4 x 2.0mm, 10/Pk Part No.: AJ0-4286

- Holder Part No.: KJ0-4282



Products used in this application:



ANALYTES:

- 1 Homocysteine
- 2 Cysteine
- 3 Cystinyl-glycine
- 4 Glutathione

