


## Metanephrine

**Column:** N/A**Dimensions:** N/A**Order No:** N/A**Elution Type:** N/A**Eluent A:** Hexane/Dichloroethane/Ethanol-TFA (55:35:10) ethanol-TFA 20:1**Gradient****Profile:****Flow Rate:** 1 mL/min**Col. Temp.:** ambient**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 280 nm (ambient) $\alpha = 1.12$ 

### ANALYTES:

**1** Metanephrine

Retention Time: 3.24 min

**2** Metanephrine

Retention Time: 19.05 min

**3** Metanephrine

Retention Time: 21.01 min

