


Prilocaine

Column: N/A
Dimensions: N/A
Order No: N/A
Elution Type: N/A
Eluent A: Hexane/Dichloroethane/Ethanol-TFA (60:35:5) ethanol-TFA 20:1
Gradient

| Step No. | Time (min) | Pct A |
|----------|------------|-------|
| 1 | 0 | 100 |

Profile:
Flow Rate: 1 mL/min
Col. Temp.: ambient
Detection: UV-Vis Abs.-Variable Wave.(UV) @ 242 nm (ambient)



ANALYTES:

- 1 Prilocaine
Retention Time: 3.15 min
- 2 Prilocaine
Retention Time: 13.16 min
- 3 Prilocaine
Retention Time: 14.57 min

