

HPLC Application

ID No.: 14979

Apple Juice analysis using Rezex RCM

Column: Rezex[™] RCM-Monosaccharide Ca+2 (8%), LC Column 300 x 7.8 mm, Ea

Dimensions: 300 x 7.8 mm ID

Order No: 00H-0130-K0

Elution Type: Isocratic

Eluent A: Water

Gradient	Step No.	Time (min)	Pct A
Profile:	1	0	100

Flow Rate: 0.6 mL/min

Col. Temp.: 80 °C


Detection: Evaporative Light Scattering @ nm (ambient)

Analyst Note: SecurityGuard[™] Guard Cartridge System extends column lifetime.

- SecurityGuard Cartridges, Carbo-Ca 4 x 3.0mm, 10/Pk Part No.: AJ0-4493


- Holder Part No.: KJ0-4282

14979



ANALYTES:

- 1 Sucrose
- 2 Glucose
- 3 Fructose
- 4 Sorbitol



Products used in this application:

