

HPLC Application

ID No.: 15009

Basic compounds, pH 7.0

Column: Gemini[®] 5 μ m C18 110 \AA , LC Column 150 x 4.6 mm, Ea

Dimensions: 150 x 4.6 mm ID

Order No: 00F-4435-E0

Elution Type: Isocratic

Eluent A: 20mM KH₂PO₄ buffer, pH 7.0

Eluent B: Acetonitrile

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	75	25

Flow Rate: 1.5 mL/min

Col. Temp.: 25 °C

Detection: UV-Vis Abs.-Variable Wave.(UV) @ 210 nm (25 °C)

Analyst Note: SecurityGuard[™] Guard Cartridge System extends column lifetime.

- SecurityGuard Cartridges, Gemini C18 4 x 3.0mm, 10/Pk Part No.: AJ0-7597


- Holder Part No.: KJ0-4282



Products used in this application:



15009



ANALYTES:

- 1** Maleic acid
Retention Time: 0.992 min
- 2** Chlorpheniramine
Retention Time: 7.874 min
- 3** Triprolidine
Retention Time: 12.585 min
- 4** Diphenhydramine
Retention Time: 15.729 min

