

# HPLC Application

ID No.: 15473

## Beta Blockers on Onyx C18 - Effect of Flow Rate

**Column:** Onyx<sup>™</sup> Monolithic C18, LC Column 100 x 4.6 mm, Ea

**Dimensions:** 100 x 4.6 mm ID

**Order No:** CH0-7643

**Elution Type:** Isocratic

**Eluent A:** 0.1% TFA in Water

**Eluent B:** Acetonitrile

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	80	20

**Flow Rate:** 1 mL/min


**Col. Temp.:** ambient

**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 220 nm (25 °C)

15473



Products used in this application:



### ANALYTES:

- 1 Atenolol
- 2 Pindolol
- 3 Metoprolol
- 4 Celiprolol
- 5 Bisoprolol

