

# HPLC Application

ID No.: 15763

## Uracil family using Synergi Fusion-RP (acetate)

**Column:** Synergi<sup>™</sup> 4  $\mu$ m Fusion-RP 80 Å, LC Column 150 x 4.6 mm, Ea

**Dimensions:** 150 x 4.6 mm ID

**Order No:** 00F-4424-E0

**Elution Type:** Gradient

**Eluent A:** 20mM Ammonium acetate, pH 4.5

**Eluent B:** Acetonitrile

<b>Gradient Profile:</b>	<b>Step No.</b>	<b>Time (min)</b>	<b>Pct A</b>	<b>Pct B</b>
	<b>1</b>	0	97	3
	<b>2</b>	10	80	20

**Flow Rate:** 1 mL/min

**Col. Temp.:** ambient

**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 260 nm (ambient)


**Analyst Note:** SecurityGuard<sup>™</sup> Guard Cartridge System extends column lifetime.

- SecurityGuard Cartridges, Fusion-RP 4 x 3.0mm, 10/Pk Part No.: AJ0-7557

- Holder Part No.: KJ0-4282



15763



### ANALYTES:

- 1 UTP
- 2 UDP
- 3 UMP
- 4 Uracil
- 5 Uridine

