

# HPLC Application

ID No.: 16491

## SFC - Polar Selectivity Mix on Luna CN 250 x 4.6

**Column:** Luna® 5 µm CN 100 Å, LC Column 250 x 4.6 mm, Ea

**Dimensions:** 250 x 4.6 mm ID

**Order No:** 00G-4255-E0

**Elution Type:** Gradient

**Eluent A:** Methanol

**Eluent B:** Carbon Dioxide

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	5	95
	2	0.1	5	95
	3	5	50	50
	4	7	50	50

**Flow Rate:** 3 mL/min

**Col. Temp.:** 40 °C


**Detection:** UV-Vis Abs.-Fixed Wave. (UV) @ 254 nm (40 °C)

**Detector Info:** Thar Discovery Series, Super Pure SFC instrument provided courtesy of Thar Technologies

**Analyst Note:** SecurityGuard™ Guard Cartridge System extends column lifetime.

- SecurityGuard Cartridges, CN 4 x 3.0mm, 10/PK Part No.: AJ0-4305

- Holder Part No.: KJ0-4282



### ANALYTES:

- 1 Benzoic acid
- 2 Caffeine
- 3 Acetaminophen



Products used in this application:

