

## Column Test Mix on Kinetex 2.6µm C18 150 x 4.6mm ID

**Column:** Kinetex® 2.6 µm C18 100 Å, LC Column 150 x 4.6 mm, Ea

**Dimensions:** 150 x 4.6 mm ID

**Order No:** 00F-4462-E0

**Elution Type:** Isocratic

**Eluent A:** water

**Eluent B:** acetonitrile

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	30	70


**Flow Rate:** 1.8 mL/min

**Col. Temp.:** ambient

**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 254 nm (ambient)  
18185



Products used in this application:



### ANALYTES:

- 1** Uracil  
Retention Time: 0.684 min
- 2** Acetophenone  
Retention Time: 1.038 min
- 3** Benzene  
Retention Time: 1.325 min
- 4** Toluene  
Retention Time: 1.601 min
- 5** Naphthalene  
Retention Time: 1.858 min

