

## SRM 870 Acquity UPLC - Kinetex 2.6µm C18 50 x 2.1mm ID (0.42mL/min)

**Column:** Kinetex® 2.6 µm C18 100 Å, LC Column 50 x 2.1 mm, Ea

**Dimensions:** 50 x 2.1 mm ID

**Order No:** 00B-4462-AN

**Elution Type:** Isocratic

**Eluent A:** 20 mM KH<sub>2</sub>PO<sub>4</sub>

**Eluent B:** Methanol

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	20	80


**Flow Rate:** 0.42 mL/min

**Col. Temp.:** ambient

**Detection:** UV-Vis Abs.-Diode Array (PDA) @ 254 nm (25 °C)



Products used in this application:



### ANALYTES:

**1** Toluene

Retention Time: 0.523 min

**2** Ethylbenzene

Retention Time: 0.605 min

**3** Quinizarine

Retention Time: 0.918 min

**4** Amitriptyline

Retention Time: 1.072 min

