

HPLC Application

ID No.: 18568

QC Test Chromatogram (Waters Acquity UPLC) - Kinetex 2.6 μ m C18 50 x 2.1mm ID

Column: Kinetex[®] 2.6 μ m C18 100 \AA , LC Column 50 x 2.1 mm, Ea

Dimensions: 50 x 2.1 mm ID

Order No: 00B-4462-AN

Elution Type: Isocratic

Eluent A: Acetonitrile


Eluent B: Water

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	50	50

Flow Rate: 0.49 mL/min

Col. Temp.: ambient

Detection: UV-Vis Abs.-Variable Wave.(UV) @ 254 nm (ambient)



Products used in this application:



ANALYTES:

1 Uracil

Retention Time: 0.307 min

2 Acetophenone

Retention Time: 0.483 min

3 Toluene

Retention Time: 0.979 min

4 Naphthalene

Retention Time: 1.309 min

