

# HPLC Application

ID No.: 18577

## Prodigy Test Mix on Kinetex 2.6 $\mu$ m C18 50 x 2.1mm ID (without KrudKatcher Ultra)

**Column:** Kinetex<sup>®</sup> 2.6  $\mu$ m C18 100 Å, LC Column 50 x 2.1 mm, Ea

**Dimensions:** 50 x 2.1 mm ID

**Order No:** 00B-4462-AN

**Elution Type:** Isocratic

**Eluent A:** Acetonitrile

**Eluent B:** Water

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	65	35


**Flow Rate:** 0.5 mL/min

**Col. Temp.:** ambient

**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 254 nm (ambient)



Products used in this application:



### ANALYTES:

**1** Uracil

Retention Time: 0.188 min

**2** Acetophenone

Retention Time: 0.395 min

**3** Toluene

Retention Time: 0.894 min

**4** Naphthalene

Retention Time: 1.227 min

