

HPLC Application

ID No.: 18577

Prodigy Test Mix on Kinetex 2.6 μ m C18 50 x 2.1mm ID (without KrudKatcher Ultra)


Column: Kinetex[®] 2.6 μ m C18 100 Å, LC Column 50 x 2.1 mm, Ea
Dimensions: 50 x 2.1 mm ID
Order No: 00B-4462-AN
Elution Type: Isocratic
Eluent A: Acetonitrile
Eluent B: Water
Gradient Profile:

Step No.	Time (min)	Pct A	Pct B
1	0	65	35

Flow Rate: 0.5 mL/min
Col. Temp.: ambient
Detection: UV-Vis Abs.-Variable Wave.(UV) @ 254 nm (ambient)



Products used in this application:



ANALYTES:

- 1** Uracil
Retention Time: 0.188 min
- 2** Acetophenone
Retention Time: 0.395 min
- 3** Toluene
Retention Time: 0.894 min
- 4** Naphthalene
Retention Time: 1.227 min

