

# HPLC Application

ID No.: 19096

## Compare BioSep-SEC-S3000(2) & -S3000 Hu Serum, MP w. Arginine

**Column:** BioSep<sup>™</sup> 5  $\mu$ m SEC-s3000 290 Å, LC Column 300 x 7.8 mm, Ea

**Dimensions:** 300 x 7.8 mm ID

**Order No:** 00H-2146-K0

**Elution Type:** Isocratic

**Eluent A:** 100mM sodium phosphate, 200mM Arginine, pH 6.8

Gradient	Step No.	Time (min)	Pct A
Profile:	1	0	100


**Flow Rate:** 1 mL/min

**Col. Temp.:** ambient

**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 280 nm (ambient)

**Analyst Note:** SecurityGuard<sup>™</sup> Guard Cartridge System extends column lifetime.


- SecurityGuard Cartridges, GFC 3000 4 x 3.0mm ID, 10/Pk Part No.: AJ0-4488
- Holder Part No.: KJ0-4282



Products used in this application:



19096



### ANALYTES:

- 1 Human Serum

