

HPLC Application

ID No.: 19454

Insulin impurity profile using Kinetex 2.6 μ m C18 150x4.6mm

Column: Kinetex[®] 2.6 μ m C18 100 Å, LC Column 150 x 4.6 mm, Ea

Dimensions: 150 x 4.6 mm ID

Order No: 00F-4462-E0

Elution Type: Gradient

Eluent A: 20mM Potassium phosphate pH 2.7

Eluent B: Acetonitrile

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	76	24
	2	30	10	90

Flow Rate: 1.2 mL/min


Col. Temp.: ambient

Detection: UV-Vis Abs.-Diode Array (PDA) @ (ambient)

19454



Products used in this application:



ANALYTES:

1 Insulin

