

# HPLC Application

ID No.: 19519

## Simvastatin by HPLC-Kinetex 2.6µm XB-C18 150x4.6mm

**Column:** Kinetex® 2.6 µm XB-C18 100 Å, LC Column 150 x 4.6 mm, Ea

**Dimensions:** 150 x 4.6 mm ID

**Order No:** 00F-4496-E0

**Elution Type:** Isocratic

**Eluent A:** 65:35 ACN:28mM sodium phosphate pH 4.5


Gradient	Step No.	Time (min)	Pct A
<b>Profile:</b>	<b>1</b>	15	100

**Flow Rate:** 1.5 mL/min

**Col. Temp.:** ambient

**Detection:** UV-Vis Abs.-Diode Array (PDA) @ 238 nm (ambient)

19519



### ANALYTES:

**1** Lovastatin

Retention Time: 3.848 min

**2** Simvastatin

Retention Time: 5.194 min



Products used in this application:

