

# HPLC Application

ID No.: 19526

## Steroids Mix using Kinetex 2.6µm C8, 50x2.1mm

**Column:** Kinetex® 2.6 µm C8 100 Å, LC Column 50 x 2.1 mm, Ea

**Dimensions:** 50 x 2.1 mm ID

**Order No:** 00B-4497-AN

**Elution Type:** Gradient

**Eluent A:** water

**Eluent B:** ACN


Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	80	20
	2	6	40	60
	3	6.01	80	20
	4	8	80	20

**Flow Rate:** 0.5 mL/min

**Col. Temp.:** ambient

**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 220 nm (ambient)

19526



### ANALYTES:

- 1** Estriol  
Retention Time: 1.002 min
- 2** Hydrocortisone  
Retention Time: 1.222 min
- 3** Cortisone  
Retention Time: 1.286 min
- 4** Cortisone-21-acetate  
Retention Time: 2.502 min
- 5** Estradiol  
Retention Time: 2.533 min
- 6** 21-Hydroxyprogesterone  
Retention Time: 2.673 min
- 7** Estrone  
Retention Time: 2.945 min
- 8** 17-Hydroxyprogesterone  
Retention Time: 2.945 min
- 9** Deoxycorticosterone acetate  
Retention Time: 3.958 min



Products used in this application:

