

HPLC Application

ID No.: 19549

Prodigy Test Mix using Kinetex 2.6µm XB-C18 50x2.1mm

Column: Kinetex® 2.6 µm XB-C18 100 Å, LC Column 50 x 2.1 mm, Ea

Dimensions: 50 x 2.1 mm ID

Order No: 00B-4496-AN

Elution Type: Isocratic

Eluent A: Water

Eluent B: ACN

Gradient

Step No.	Time (min)	Pct A	Pct B
1	0	50	50


Profile:

Flow Rate: 0.5 mL/min

Col. Temp.: ambient

Detection: UV-Vis Abs.-Variable Wave.(UV) @ 254 nm (ambient)

19549



ANALYTES:

- 1** Uracil
Retention Time: 0.219 min
- 2** Acetophenone
Retention Time: 0.441 min
- 3** Toluene
Retention Time: 0.966 min
- 4** Naphthalene
Retention Time: 1.309 min



Products used in this application:

