

# HPLC Application

ID No.: 19554

## Prodigy Test Mix using Kinetex 2.6µm C18 50x2.1mm

**Column:** Kinetex® 2.6 µm C18 100 Å, LC Column 50 x 2.1 mm, Ea

**Dimensions:** 50 x 2.1 mm ID

**Order No:** 00B-4462-AN

**Elution Type:** Isocratic

**Eluent A:** Water

**Eluent B:** ACN


Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	50	50

**Flow Rate:** 0.5 mL/min

**Col. Temp.:** ambient

**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 254 nm (ambient)

19554



### ANALYTES:

**1** Uracil

Retention Time: 0.219 min

**2** Acetophenone

Retention Time: 0.437 min

**3** Toluene

Retention Time: 0.896 min

**4** Naphthalene

Retention Time: 1.191 min



Products used in this application:

