

HPLC Application

ID No.: 19555

Prodigy Test Mix using Kinetex 2.6µm C8 50x2.1mm ID

Column: Kinetex® 2.6 µm C8 100 Å, LC Column 50 x 2.1 mm, Ea

Dimensions: 50 x 2.1 mm ID

Order No: 00B-4497-AN

Elution Type: Isocratic

Eluent A: Water

Eluent B: ACN


Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	50	50

Flow Rate: 0.5 mL/min

Col. Temp.: ambient

Detection: UV-Vis Abs.-Variable Wave.(UV) @ 254 nm (ambient)

19555



ANALYTES:

- 1** Uracil
Retention Time: 0.216 min
- 2** Acetophenone
Retention Time: 0.402 min
- 3** Toluene
Retention Time: 0.762 min
- 4** Naphthalene
Retention Time: 0.959 min



Products used in this application:

