
HPLC Application

Separation of Testosterone and Epitestosterone by LC/MS/MS using Strata-X-A and Kinetex 1.7 µm C18

Column:	Kinetex® 1.	7 μm C18 100 Å, L	.C Column 30	x 2.1 mm, Ea					
Dimensions:	30 x 2.1 mm	ו ID							
Order No:	00A-4475-A	N		Kinetex [®]					
Elution Type:	Gradient			Ultra-High Performance on Any LC System					
Eluent A:	0.1% Formio	c Acid +1 mM Amm	n Formate in						
Eluent B:	0.1% Formio	c Acid +1 mM Amm							
Gradient	Step No.	Time (min)	Pct A	Pct B					
Profile:	1	0	90	10					
	2	2.5	10	90	Products used in this application:				
	3	3.5	10	90	Products used in this application.				
	4	3.6	90	10					
Flow Rate:	0.4 mL/min			(KINETEX.					
Col. Temp.:	ambient								
Detection:	Mass Spectrometer (MS) @ amu (ambient)								
Detector Info: 19761	<a <br="" target="_blank">href="bttps://sciex.com/products/mass-spectrometers?utm_campaign=2019%20application%20search&utm_source=phenomenex&utm_medium=referral">SCIEX<,								



ANALYTES:

- 1 Testosterone Retention Time: 2.62 min
- 2 Testosterone-d3 Retention Time: 2.61 min
- 3 Epitestosterone

Retention Time: 2.77 min

©2025 Phenomenex Inc. All rights reserved.

Phenomenex products are available worldwide.

www.phenomenex.com

For more information contact your Phenomenex Representative at support@phenomenex.com

phenomenex

Sample Preparation Details

for HPLC Application ID No.: 19761



Separation of Testosterone and Epitestosterone by LC/MS/MS using Strata-X-A and Kinetex 1.7 µm C18

PRODUCT DESCRIPTION:

Strata^m-X-A 33 µm Polymeric Strong Anion, 30 mg / 3 mL, Tubes , 50/Pk

Order No.: 8B-S123-TBJ

SOLID PHASE EXTRACTION (SPE) PRODCEDURE:

Note: The solvent volumes shown below are for a 30 mg bed mass.

The solvent volumes will need to be adjusted for a smaller or larger bed mass.

Condition:

Load:

Wash:

Dry:

Dry for 5 min under high vacuum

Elute:

Final Prep and Analysis:

Following evaporation of elution solvent @ 50-55 C under gentle nitrogen stream; Add 50 uL 25% hydroxylamine solution and heat at 60-65 C for 5-10 min, then add 200 uL 5%

Inject: 0 µL on HPLC Mass Spectrometer (MS) @ amu (ambient)

ANALYTES:	Spiked Conc.	Log P	рКа	% Rec	%RSC
	(ng/mL)				(n=0)
1 Testosterone	100				
2 Testosterone-d3	0.5				
3 Epitestosterone	0				

Note: This method is designed as a convenient starting point for further investigation and can be tailored to meet your extraction goals. Call your local Phenomenex Representative for assistance in method development and optimization techniques.

©2025 Phenomenex Inc. All rights reserved.

Phenomenex products are available worldwide.

www.phenomenex.com

For more information contact your Phenomenex Representative at support@phenomenex.com