## **HPLC Application**

ID No.: 19762



## Testosterone from Female Human Plasma by LC/MS/MS using Strata-X-A and Kinetex 1.7 µm C18

Kinetex® 1.7 μm C18 100 Å, LC Column 30 x 2.1 mm, Ea

30 x 2.1 mm ID **Dimensions:** Order No: 00A-4475-AN Elution Type: Gradient

Eluent A: 0.1% Formic Acid +1 mM Amm Formate in Water

Fluent B:	0.1% Formic Acid +1 mm Amm Formate in ACN			
Gradient	Step No.	Time (min)	Pct A	Pct B
Profile:	1	0	90	10
	2	2.5	10	90
	3	3.5	10	90
	4	3.6	90	10

Kinetex<sup>®</sup>

Products used in this application:




Flow Rate: 0.4 mL/min Col. Temp.: ambient

**Detection:** Mass Spectrometer (MS) @ amu (ambient)

**Detector Info:** <a target="\_blank"

cts/mass-spectrometers?utm\_campaign=2019%20application%20search&utm\_source=phenomenex&utm\_medium=referral">SCIEX<



### **ANALYTES:**

Testosterone

Retention Time: 2.62 min 2 Testosterone-d3 Retention Time: 2.61 min

@2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at support@phenomenex.com



# **Sample Preparation Details**

for HPLC Application ID No.: 19762



### Testosterone from Female Human Plasma by LC/MS/MS using Strata-X-A and Kinetex 1.7 µm C18

### **PRODUCT DESCRIPTION:**

Strata<sup>™</sup>-X-A 33 µm Polymeric Strong Anion, 30 mg / 3 mL, Tubes , 50/Pk

Order No.: 8B-S123-TBJ

### **SOLID PHASE EXTRACTION (SPE) PRODCEDURE:**

Note: The solvent volumes shown below are for a 30 mg bed mass.

The solvent volumes will need to be adjusted for a smaller or larger bed mass.

Condition:	_			
.oad:	_			
Wash:	_			
Ory:				
Dry for 5 min under high vacuum	_			
Elute:				
Final Prep and Analysis:	_			
Following evaporation of elution solvent @ 50-55 C under gentle nitrogen stream; Add 50 uL 25% hydroxylamine solution and heat at 60-65 C for 5-10 min, then add 200 uL 5%				
Inject: 0 µL on HPLC Mass Spectrometer (MS) @ amu (ambient)				

Spiked Conc. %RSC **ANALYTES:** Log P pKa % Rec (ng/mL) (n=0)1 Testosterone 0 2 Testosterone-d3 0.5

This method is designed as a convenient starting point for further investigation and can be tailored to meet your extraction goals. Call your local Phenomenex Representative for assistance in method development and optimization techniques.

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at support@phenomenex.com

