

HPLC Application

ID No.: 19816

Angiotensin 1 on Aeris WIDEPORÉ 3.6u XB-C18

Column: Aeris[™] 3.6 µm WIDEPORÉ XB-C18 200 Å, LC Column 150 x 4.6 mm, Ea

Dimensions: 150 x 4.6 mm ID

Order No: 00F-4482-E0

Elution Type: Gradient

Eluent A: 0.1% Formic in Water

Eluent B: Acetonitrile

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	97	3
	2	3	97	3
	3	23	35	65

Flow Rate: 1.2 mL/min

Col. Temp.: 40 °C


Detection: UV-Vis Abs.-Variable Wave.(UV) @ 214 nm (ambient)

Analyst Note: SecurityGuard[™] Guard Cartridge System extends column lifetime.


- SecurityGuard ULTRA Cartridges, UHPLC WIDEPORÉ C18 for 4.6mm ID, 3/Pk Part No.: AJ0-8769

- Holder Part No.: AJ0-9000

19816



Products used in this application:



ANALYTES:

1 Angiotensin

Retention Time: 9.942 min

