

HPLC Application

ID No.: 19831

LCMS Bleed Chart for Aeris WIDEPORE

Column: Aeris™ 3.6 µm WIDEPORE XB-C18 200 Å, LC Column 50 x 2.1 mm, Ea

Dimensions: 50 x 2.1 mm ID

Order No: 00B-4482-AN

Elution Type: Gradient

Eluent A: Water with 0.1% Formic acid

Eluent B: Acetonitrile with 0.1% Formic acid

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	95	5
	2	2.5	5	95
	3	3	5	95
	4	3.1	95	5

Flow Rate: 0.5 mL/min

Col. Temp.: ambient


Detection: Mass Spectrometer (MS) @ amu (450 °C)

Detector Info: <a target="_blank"


Analyst Note: href="https://sciex.com/products/mass-spectrometers?utm_campaign=2019%20application%20search&utm_source=phenomenex&utm_medium=referral">SCIEX
SecurityGuard™ Guard Cartridge System extends column lifetime.

- SecurityGuard ULTRA Cartridges, UHPLC WIDEPORE C18 for 2.1mm ID, 3/Pk Part No.: AJ0-8783

- Holder Part No.: AJ0-9000



Products used in this application:



ANALYTES:

- 1 bleed profile

