

## LCMS Bleed Chart for Aeris WIDEPORE

**Column:** Aeris™ 3.6 µm WIDEPORE XB-C18 200 Å, LC Column 50 x 2.1 mm, Ea

**Dimensions:** 50 x 2.1 mm ID

**Order No:** 00B-4482-AN

**Elution Type:** Gradient

**Eluent A:** Water with 0.1% Formic acid

**Eluent B:** Acetonitrile with 0.1% Formic acid

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	95	5
	2	2.5	5	95
	3	3	5	95
	4	3.1	95	5

**Flow Rate:** 0.5 mL/min

**Col. Temp.:** ambient


**Detection:** Mass Spectrometer (MS) @ amu (450 °C)

**Detector Info:** <a target="\_blank"


**Analyst Note:** href="https://sciex.com/products/mass-spectrometers?utm\_campaign=2019%20application%20search&utm\_source=phenomenex&utm\_medium=referral">SCIEX<br/>SecurityGuard™ Guard Cartridge System extends column lifetime.

- SecurityGuard ULTRA Cartridges, UHPLC WIDEPORE C18 for 2.1mm ID, 3/Pk Part No.: AJ0-8783

- Holder Part No.: AJ0-9000



Products used in this application:



### ANALYTES:

- 1 bleed profile

