

# HPLC Application

ID No.: 20053

## Pharma ABN Test 15 Mix on Waters Acquity UPLC using Kinetex 2.6u C18 100x2.1mm @ 0.60mL/min

**Column:** Kinetex® 2.6 µm C18 100 Å, LC Column 100 x 2.1 mm, Ea

**Dimensions:** 100 x 2.1 mm ID

**Order No:** 00D-4462-AN

**Elution Type:** Gradient

**Eluent A:** 0.1% Formic Acid in Water

**Eluent B:** 0.1% Formic Acid in Acetonitrile

<b>Gradient Profile:</b>	<b>Step No.</b>	<b>Time (min)</b>	<b>Pct A</b>	<b>Pct B</b>
	<b>1</b>	0	95	5
	<b>2</b>	0.2	95	5
	<b>3</b>	5	5	95
	<b>4</b>	6	5	95
	<b>5</b>	6.1	95	5
	<b>6</b>	7	95	5

**Flow Rate:** 0.6 mL/min

**Col. Temp.:** ambient

**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 254 nm (ambient)


**Analyst Note:** SecurityGuard™ ULTRA Guard Cartridge System extends column lifetime.

- SecurityGuard ULTRA Cartridges, UHPLC C18 for 2.1mm ID Columns, 3/Pk Part No.: AJ0-8782

- Holder Part No.: AJ0-9000



Products used in this application:



## Pharma ABN Test 15 Mix on Waters Acquity UPLC using Kinetex 2.6u C18 100x2.1mm @ 0.60mL/min

### ANALYTES:

- 1 Pyridine
- 2 Acetaminophen
- 3 Pindolol
- 4 Sulfathiazole
- 5 Acebutolol
- 6 Chlorpheniramine
- 7 Benzyl alcohol
- 8 Phenol
- 9 Triprolidine
- 10 3-Methyl-4-nitrobenzoic acid
- 11 Nortriptyline
- 12 Prednisolone
- 13 Diflunisal
- 14 2-Hydroxy-5-methylbenzaldehyde
- 15 Hexanophenone

