

## BioSep-SEC-S 4000 PMQC Standard Low Salt Buffer

**Column:** BioSep<sup>™</sup> 5  $\mu$ m SEC-s4000 500 Å, LC Column 300 x 7.8 mm, Ea

**Dimensions:** 300 x 7.8 mm ID

**Order No:** 00H-2147-K0

**Elution Type:** Isocratic

**Eluent A:** 100 mM Sodium Phosphate in Water pH 6.8 with 0.025% Na<sub>3</sub>


Gradient	Step No.	Time (min)	Pct A
<b>Profile:</b>	<b>1</b>	0	100

**Flow Rate:** 1 mL/min

**Col. Temp.:** ambient


**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 280 nm (ambient)

20605



### ANALYTES:

- 1** Thyroglobulin  
Retention Time: 8 min
- 2** IgA  
Retention Time: 8.7 min
- 3** IgG  
Retention Time: 9.4 min
- 4** Ovalbumin  
Retention Time: 10.4 min
- 5** Myoglobin  
Retention Time: 10.9 min
- 6** Uridine  
Retention Time: 11.9 min



Products used in this application:

