

OPA Derivatization of Tyramine and Histamine using Kinetex 2.6u C18


Column: Kinetex® 2.6 µm C18 100 Å, LC Column 100 x 4.6 mm, Ea
Dimensions: 100 x 4.6 mm ID
Order No: 00D-4462-E0
Elution Type: Gradient
Eluent A: 20mM Sodium Acetate pH 7.2 w/0.03% Sodium Azide (pre-filtered 0.2µm)
Eluent B: 50:50 Acetonitrile/Methanol

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	97	3
	2	20	20	80
	3	20.5	97	3
	4	23	97	3

Flow Rate: 1.2 mL/min
Col. Temp.: 30 °C
Detection: UV-Vis Abs.-Variable Wave.(UV) @ 338 nm (ambient)

- Detector Info:** Offline Amino Acid derivatization:
1. In to a amber vial dispense 250uL of 20mM Sodium Acetate pH 12.5
 2. Add 100uL of Sample (ex: Wine sample, Standards)
 3. Mix for 30 seconds
 4. Add 50uL of OPA derivatization agent (OPA: Part # 5061-3335)
 5. Mix vigorously for 30 seconds
 6. Add 2.0mL of MilliQ Water
 7. Mix for 30 seconds
 8. Inject 10uL of sample on to system

20873



Products used in this application:



OPA Derivatization of Tyramine and Histamine using Kinetex 2.6u C18

ANALYTES:

- 1 Histamine
- 2 Tyramine

