

## Human Plasma Vitamin C on Kinetex 5u XB-C18, 150x4.6mm Wavelength Study

**Column:** Kinetex® 5µm XB-C18 100 Å, LC Column 150 x 4.6 mm, Ea

**Dimensions:** 150 x 4.6 mm ID

**Order No:** 00F-4605-E0

**Elution Type:** Gradient

**Eluent A:** 0.1% formic acid

**Eluent B:** Acetonitrile 100%

| Gradient Profile: | Step No. | Time (min) | Pct A | Pct B |
|-------------------|----------|------------|-------|-------|
|                   | 1        | 0          | 100   | 0     |
|                   | 2        | 3.5        | 100   | 0     |
|                   | 3        | 3.6        | 0     | 100   |
|                   | 4        | 5          | 0     | 100   |
|                   | 5        | 5.1        | 100   | 0     |
|                   | 6        | 7          | 100   | 0     |

**Flow Rate:** 800 µL/min

**Col. Temp.:** ambient

**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 245 nm (ambient)

**Analyst Note:** Guard column and cartridges were used for human plasma vitamin C analysis:

SecurityGuard Ultra Cartridges: Cat log No. AJO-8768.


SecurityGuard Ultra Cartridge holder: Cat log No. AJO-9000.



Products used in this application:



21442



### ANALYTES:

1 Vitamin C

Retention Time: 2.205 min



# Sample Preparation Details

for HPLC Application ID No.: 21442

## Human Plasma Vitamin C on Kinetex 5u XB-C18, 150x4.6mm Wavelength Study

### PRODUCT DESCRIPTION:

Impact<sup>™</sup> Protein Precipitation, 2mL Square Well Filter Plate, 2/Pk

Order No.: CE0-7565

### SOLID PHASE EXTRACTION (SPE) PROCEDURE:

**Note:** The solvent volumes shown below are for a Proprietary bed mass.

The solvent volumes will need to be adjusted for a smaller or larger bed mass.

#### Condition:

---

#### Load:

---

#### Wash:

---

#### Dry:

---

#### Elute:

---

#### Final Prep and Analysis:

---

Inject: 1 µL on HPLC UV-Vis Abs.-Variable Wave.(UV) @ 245 nm (ambient)

| <b>ANALYTES:</b> | <b>Spiked Conc.<br/>(ng/mL)</b> | <b>Log P</b> | <b>pKa</b> | <b>% Rec</b> | <b>%RSC<br/>(n=0)</b> |
|------------------|---------------------------------|--------------|------------|--------------|-----------------------|
| 1 Vitamin C      | 50                              |              |            |              |                       |

**Note:** This method is designed as a convenient starting point for further investigation and can be tailored to meet your extraction goals. Call your local Phenomenex Representative for assistance in method development and optimization techniques.

©2025 Phenomenex Inc. All rights reserved.

For more information contact your Phenomenex Representative at support@phenomenex.com



Phenomenex products are available worldwide.

[www.phenomenex.com](http://www.phenomenex.com)

[support@phenomenex.com](mailto:support@phenomenex.com)