

HPLC Application

ID No.: 21669

Bitertanol on Lux 5µm Cellulose-4 in NP

Column: Lux[®] 5 µm Cellulose-4, LC Column 250 x 4.6 mm, Ea

Dimensions: 250 x 4.6 mm ID

Order No: 00G-4491-E0

Elution Type: Isocratic

Eluent A: Hexane / IPA (80:20) DEA (0.1%)

Gradient	Step No.	Time (min)	Pct A
Profile:	1	0	100


Flow Rate: 1 mL/min

Col. Temp.: ambient

Detection: UV-Vis Abs.-Variable Wave.(UV) @ 220 nm (ambient)
21669



Products used in this application:



ANALYTES:

- 1** Bitertanol
Retention Time: 6.614 min
- 2** Bitertanol
Retention Time: 8.177 min