HPLC Application

Extraction of Phenylbutazone from Beef Using Strata-X-A on Kinetex 2.6u XB-C18 50x2.1 by LC/MS/MS

Column:	Kinetex® 2	.6 μm XB-C18 100							
Dimensions:	50 x 2.1 mi	m ID							
Order No:	00B-4496-A	AN	KINETEX [®] Ultra-High Performan	ICE					
Elution Type:	Gradient on Any LC Sy								
Eluent A:	0.1% Form								
Eluent B:	0.1% Form	ic Acid in Methanol							
Gradient	Step No.	Time (min)	Pct A	Pct B					
Profile:	1	0	50	50					
	2	0.5	50	50		Products used in this applicatio	on:		
	3	2	5	95					
	4	3.5	5	95					
	5	3.51	50	50			Κ.		
	6	5	50	50		IN IN			
Flow Rate:	0.45 mL/mi	in							
Col. Temp.:	45 °C								
Detection:	Mass Spect	rometer (MS) @ an	nu (ambient))					
Detector Info:	<a https:="" s<="" target="_bl
href=" th=""><th colspan="8"><a <br="" target="_blank">href="https://sciex.com/products/mass-spectrometers?utm_campaign=2019%20application%20search&utm_source=phenomenex&utm_medium=referral">SCIE"</th>	<a <br="" target="_blank">href="https://sciex.com/products/mass-spectrometers?utm_campaign=2019%20application%20search&utm_source=phenomenex&utm_medium=referral">SCIE"							
Analyst Note:	Mass spec cond	utions: Turketer Course i terioch							
	API 4000, ESI	c oo	ion;						
	CAD:	6.00							
	CUR:	20.00							
	GSI:	50.00							
	GS2:	50.00							
	15:	5500.00							
	IEM:	600.00							
	ine:								
	DP	50.00							
	EP 21872	10.00							
·	21072								
1.9e4					Max. 1.3e4 cps.				
1.0e4			٨						
0.0	AM		Ann						
C	0.2 0.4 0.6 0.8 1.0	1.2 1.4 1.6 1.8 2.0 2.2 2.4	4 2.6 2.8 3.0 3.2	3.4 3.6 3.8 4.0 4.2	4.4 4.6 4.8 min				
1.3e4	1	1			Max. 1.3e4 cps.				
1.004		309.2→120.0							
0.0	JWh				· · · · · · · · · · · · · · · · · · ·				
0	0.2 0.4 0.6 0.8 1.0	1.2 1.4 1.6 1.8 2.0 2.2 2.4	4 2.6 2.8 3.0 3.2	3.4 3.6 3.8 4.0 4.2	4.4 4.6 4.8 min				
5083		1		MP and matrix	Max. 5083.3 cps.				
2000 -		MRM Channel 2 309.2→160.0			M				
0	0.2 0.4 0.6 0.8 1.0	1.2 1.4 1.6 1.8 2.0 2.2 2.4	2.6 2.8 3.0 3.2	3.4 3.6 3.8 4.0 4.2	4.4 4.6 4.8 min				
1001			-		Max. 1.3e4 cps.				
1.3e4 1.0e4 5000.0	N	MRM Channel 3 309.2→188.0							
0 1 0	0.2 0.4 0.6 0.8 1.0	1.2 1.4 1.6 1.8 2.0 2.2 2.4	4 2.6 2.8 3.0 3.2	3.4 3.6 3.8 4.0 4.2	4.4 4.6 4.8 min				

©2025 Phenomenex Inc. All rights reserved.

Phenomenex products are available worldwide.

www.phenomenex.com

For more information contact your Phenomenex Representative at support@phenomenex.com

• phenomenex ...breaking with tradition**

HPLC Application



Extraction of Phenylbutazone from Beef Using Strata-X-A on Kinetex 2.6u XB-C18 50x2.1 by LC/MS/MS

ANALYTES:

Phenylbutazone 1

©2025 Phenomenex Inc. All rights reserved.

Phenomenex products are available worldwide.

www.phenomenex.com

For more information contact your Phenomenex Representative at support@phenomenex.com

for HPLC Application ID No.: 21872



Extraction of Phenylbutazone from Beef Using Strata-X-A on Kinetex 2.6u XB-C18 50x2.1 by LC/MS/MS

PRODUCT DESCRIPTION:

Strata[™]-X-A 33 µm Polymeric Strong Anion, 100 mg / 3 mL, Tubes , 50/Pk

Order No.: 8B-S123-EBJ

SOLID PHASE EXTRACTION (SPE) PRODCEDURE:

Note: The solvent volumes shown below are for a 100 mg bed mass.

The solvent volumes will need to be adjusted for a smaller or larger bed mass.

Condition:

Load:

Sample Pretreatment:

- 1. In an Erleneyer flask (or similar), combine 2 ±0.1 g sample (well-homogenized ground beef or pork sausage) with 2 mL 0.1N NaOH and 50 uL of 4 ug/mL Int Std (phenylbutazone-D10) solution
- 2. Cap the flasks and mix vigorously for at least 10 sec
- 3. Add 8 mL 100% MeOH to the flask and cap the vessel
- 4. Place the flasks securely on a lab shaker and shake the vessels for 55-60 min at medium-high to high setting
- 5. Transfer the contents of the flask into a 15 mL conical tube and centrifuge them for 10 min @ 4000-4500 rpm
- 6. Remove 2 mL of supernatant and combine with 2 mL DI water into a glass tube and mix

7. Proceed to SPE method

Wash:

Dry:

5 mins at 10" of Hg

Elute:

Final Prep and Analysis:

Extraction Procedure:

1. Condition a Strata-X-A, 100 mg/6 mL with 3 mL 100% MeOH followed by 3 mL DI water

Inject: 5 µL on HPLC Mass Spectrometer (MS) @ amu (ambient)

ANALYTES:	Spiked Conc.	Log P	рКа	% Rec	%RSC
	(ng/mL)				(n=0)
1 Phenylbutazone	25			101	

Note: This method is designed as a convenient starting point for further investigation and can be tailored to meet your extraction goals. Call your local Phenomenex Representative for assistance in method development and optimization techniques.

Phenomenex products are available worldwide.

www.phenomenex.com

For more information contact your Phenomenex Representative at support@phenomenex.com