

HPLC Application

ID No.: 2439

Amiodarone

Column: N/A

Dimensions: N/A

Order No: N/A

Elution Type: N/A

Eluent A: 0.1% TFA in water

Eluent B: 0.1% TFA in acetonitrile

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	95	5
	2	15	35	65
	3	30	5	95

Flow Rate: 0 mL/min

Col. Temp.: ambient

Detection: UV-Vis Abs.-Variable Wave.(UV) @ 240 nm (ambient)

Analyst Note: SecurityGuard[™] Guard Cartridge System extends column lifetime.

- SecurityGuard Cartridges, C8 4 x 3.0mm, 10/Pk Part No.: AJ0-4290

- Holder Part No.: KJ0-4282

No Chromatogram Available

ANALYTES:

- 1 Amiodarone I,II,III,VI,VII,VIII,X,XI,XII,XIII,XIV

