

# HPLC Application

ID No.: 2439

## Amiodarone

**Column:** N/A

**Dimensions:** N/A

**Order No:** N/A

**Elution Type:** N/A

**Eluent A:** 0.1% TFA in water

**Eluent B:** 0.1% TFA in acetonitrile

<b>Gradient Profile:</b>	<b>Step No.</b>	<b>Time (min)</b>	<b>Pct A</b>	<b>Pct B</b>
	<b>1</b>	0	95	5
	<b>2</b>	15	35	65
	<b>3</b>	30	5	95

**Flow Rate:** 0 mL/min

**Col. Temp.:** ambient

**Detection:** UV-Vis Abs.-Variable Wave.(UV) @ 240 nm (ambient)

**Analyst Note:** SecurityGuard<sup>™</sup> Guard Cartridge System extends column lifetime.

- SecurityGuard Cartridges, C8 4 x 3.0mm, 10/Pk Part No.: AJ0-4290

- Holder Part No.: KJ0-4282

No Chromatogram Available

### ANALYTES:

- 1 Amiodarone I,II,III,VI,VII,VIII,X,XI,XII,XIII,XIV

