

HPLC Application

ID No.: 24665

PFASs from Milk, Eggs, Butter, Cheese and fish using Luna[®] Omega 1.6 μm PS C18 (0.5 ppb)

Column: Luna Omega 1.6u PS C18 100A, LC Column 100 x 2.1 mm, Ea

Dimensions: 100 x 2.1 mm ID

Order No: 00D-4752-AN

Elution Type: Gradient

Eluent A: 5 mM Ammonium Acetate/Water

Eluent B: Acetonitrile

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	60	40
	2	0.5	60	40
	3	3	10	90
	4	3.1	0	100
	5	4	0	100

Flow Rate: 0.55 mL/min


Col. Temp.: 40 °C

Detection: Tandem Mass Spec (MS-MS) @ (ambient)

Detector Info: Agilent 6460 QQQ



Products used in this application:



PFASs from Milk, Eggs, Butter, Cheese and fish using Luna® Omega 1.6 µm PS C18 (0.5 ppb)

ANALYTES:

- 1 PFBA
- 2 PFPeA
- 3 PFBS
- 4 PFHxA
- 5 PFHpA
- 6 PFHxS
- 7 6:2 FTS
- 8 PFOA
- 9 PFHpS
- 10 PFOS
- 11 PFNA
- 12 FOSA
- 13 Me-FOSE
- 14 8:2FTS
- 15 Me-FOSA
- 16 PFDA
- 17 Et-FOSE
- 18 Et-FOSA
- 19 PFDS
- 20 PFUdA
- 21 PFDoA
- 22 PFTTrDA
- 23 PFTeDA

