

# HPLC Application

ID No.: 25102

## SecurityLINK Zero-Dead Volume Connection vs Incorrectly Made Connection on Luna Omega C18

**Column:** Luna Omega 1.6  $\mu$ m C18 100  $\text{\AA}$ , LC Column 50 x 2.1 mm, Ea

**Dimensions:** 50 x 2.1 mm ID

**Order No:** 00B-4742-AN

**Elution Type:** Isocratic

**Eluent A:** Water

**Eluent B:** Acetonitrile

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	2	0	35	65

**Flow Rate:** 0.5 mL/min


**Col. Temp.:** 25 °C

**Detection:** LC/UV (DAD, PDA) @ 254.000000000 nm (nanometers) (ambient)

**Analyst Note:** ACN:Water (75:25) AL0-3045



Products used in this application:



### ANALYTES:

