

HPLC Application

ID No.: 25102

SecurityLINK Zero-Dead Volume Connection vs Incorrectly Made Connection on Luna Omega C18

Column: Luna Omega 1.6 μ m C18 100 Å, LC Column 50 x 2.1 mm, Ea

Dimensions: 50 x 2.1 mm ID

Order No: 00B-4742-AN

Elution Type: Isocratic

Eluent A: Water

Eluent B: Acetonitrile

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	2	0	35	65

Flow Rate: 0.5 mL/min


Col. Temp.: 25 °C

Detection: LC/UV (DAD, PDA) @ 254.000000000 nm (nanometers) (ambient)

Analyst Note: ACN:Water (75:25) AL0-3045



Products used in this application:



ANALYTES:

