

## EPA Method 624 Calibration Mix

**Column:** Zebtron™ ZB-624, GC Cap. Column 60 m x 0.25 mm x 1.40 μm, Ea

**Phase:** 6% Cyanopropylphenyl 94% Dimethylpolysiloxane

**Dimensions:** 60 meters x 0.25 mm x 1.4 μm

**Order No:** 7KG-G005-27

**Oven Profile:** 40°C (hold 4 min) to 160°C @ 5°C/min


**Carrier Gas:** Constant Flow Hydrogen, 1.03 mL/min

**Injection:** Split 100:1 1 μL @ 250°C

**Detection:** Flame Ionization (FID) (255°C)



Products used in this application:



## EPA Method 624 Calibration Mix

### ANALYTES:

- 1 1,1-Dichloroethene
- 2 Methylene chloride
- 3 1,1-Dichloroethane
- 4 Carbon tetrachloride
- 5 Benzene
- 6 Trichloroethene
- 7 1,2-Dichloropropane
- 8 2-Chlorethyl vinyl ether
- 9 1,1,2-Trichloroethane
- 10 Tetrachloroethene
- 11 Dibromochloromethane
- 12 Chlorobenzene

