

HPLC Application

ID No.: 5456

QC Test

Column: Prodigy[™] 5 µm C8 150 Å, LC Column 150 x 4.6 mm, Ea

Dimensions: 150 x 4.6 mm ID

Order No: 00F-3301-E0

Elution Type: Isocratic

Eluent A: Acetonitrile

Eluent B: Water

Gradient Profile:	Step No.	Time (min)	Pct A	Pct B
	1	0	65	35

Flow Rate: 1 mL/min

Col. Temp.: ambient

Detection: Unknown @ nm (ambient)

Analyst Note: SecurityGuard[™] Guard Cartridge System extends column lifetime.

- SecurityGuard Cartridges, C8 4 x 3.0mm, 10/Pk Part No.: AJ0-4290


- Holder Part No.: KJ0-4282



Products used in this application:

Prodigy[™]

5456



ANALYTES:

- 1 Acetophenone
- 2 Benzene
- 3 Toluene
- 4 Naphthalene

